

Certificate of Authenticity

IPS Empress Esthetic

The dental restorations used for your case were constructed of FDA approved 510k materials and hand crafted in an ISO 13485:2003 certified dental laboratory.
This is your assurance of quality and safety in materials and construction.

Item	Product Name	Manufacturer	Country of Origin	FDA 510k	Elements & Compounds
Esthetic Core	IPS Empress Esthetic	Ivoclar	Germany	K982616	>98% SiO ₂ , BaO, Al ₂ O ₃ , CaO, CeO ₂ , Na ₂ O, K ₂ O, B ₂ O ₃ <2% TiO ₂ and pigments
Veneer Porcelain	IPS Empress Esthetic Veneer	Ivoclar	Germany	K913372	SiO ₂ > 55%, Other: Al ₂ O ₃ , K ₂ O, Na ₂ O, CaO, B ₂ O ₃ , TiO ₂ , CeO ₂ and pigments



Heikas Martirosian

Heikas Martirosian, CDT
Technical Director

